



National Accreditation Board for  
Testing and Calibration Laboratories

**CERTIFICATE OF ACCREDITATION**

**SBEM PVT LTD , SBEM FLOW LABORATORY**

has been assessed and accredited in accordance with the standard

**ISO/IEC 17025:2017**

**"General Requirements for the Competence of Testing &  
Calibration Laboratories"**

for its facilities at

UNIT NO 9, 691/A/2, BIBWEWADI INDUSTRIAL ESTATE, PUNE-SATARA ROAD, PUNE, MAHARASHTRA,  
INDIA

in the field of

**CALIBRATION**

Certificate Number: CC-2440

Issue Date: 05/08/2022

Valid Until: 04/08/2024

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.  
(To see the scope of accreditation of this laboratory, you may also visit NABL website [www.nabl-india.org](http://www.nabl-india.org))

Name of Legal Identity : SBEM PVT LTD

Signed for and on behalf of NABL



N. Venkateswaran  
Chief Executive Officer



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

SBEM PVT LTD , SBEM FLOW LABORATORY, UNIT NO 9, 691/A/2, BIBWEWADI INDUSTRIAL ESTATE, PUNE-SATARA ROAD, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2440

**Page No**

1 of 2

**Validity**

05/08/2022 to 04/08/2024

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
LOCATION 1-UNIT NO 9, 691/A/2, BIBWEWADI INDUSTRIAL ESTATE, PUNE-SATARA ROAD, PUNE, MAHARASHTRA, INDIA Permanent Facility					
1	FLUID FLOW-FLOW MEASURING DEVICES	Liquid Mass /Flow Measuring Devices	Using 3000 kg weighing system as per ISO 4185 By Weighing Method	1000 kg to 3000 kg	0.05%
2	FLUID FLOW-FLOW MEASURING DEVICES	Liquid Mass Flow Rate / Flow Measuring Devices	Using 3000 kg weighing system as per ISO 4185 By Weighing Method	1.5 t/hr to 220 t/hr	0.17%
3	FLUID FLOW-FLOW MEASURING DEVICES	Liquid Volume /Flow Measuring Devices	Using 3000 kg weighing system as per ISO 4185 By Weighing Method	1000L to 3000L	0.05%
4	FLUID FLOW-FLOW MEASURING DEVICES	Liquid Volume Flow Rate (Water) / Flow measuring devices	Using 3000kg Weighing System as per ISO 4185 By Weighing Method	1.5 m <sup>3</sup> /hr to 220 m <sup>3</sup> /hr	0.5%



# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :**

SBEM PVT LTD , SBEM FLOW LABORATORY, UNIT NO 9, 691/A/2, BIBWEWADI INDUSTRIAL ESTATE, PUNE-SATARA ROAD, PUNE, MAHARASHTRA, INDIA

**Accreditation Standard**

ISO/IEC 17025:2017

**Certificate Number**

CC-2440

**Page No**

2 of 2

**Validity**

05/08/2022 to 04/08/2024

**Last Amended on**

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
LOCATION 2 - GAT NO. 326, INDUSTRIAL SHED NO E, SHRIRAMNAGAR, GAUD DARA ROAD, KHEDSHIVAPUR, PUNE, MAHARASHTRA, INDIA Permanent Facility					
1	FLUID FLOW-FLOW MEASURING DEVICES	Liquid Volume Flow Rate (Water)	Using Electromagnetic Flowmeter by Comparison Method (Comparative Method)	500 m <sup>3</sup> /hr to 3500 m <sup>3</sup> /hr	0.2%
2	FLUID FLOW-FLOW MEASURING DEVICES	Liquid Volume Measured Quantity (Water)	Using Electromagnetic Flowmeter by Comparison Method (Comparative Method)	7 m <sup>3</sup> to 35 m <sup>3</sup>	0.2%

\* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.